



中国国家强制性产品认证证书



SITIAS
Worldwide Access

证书编号: 2020322315000498

认证委托人名称: 浙江伦特机电有限公司
认证委托人地址: 浙江省温州市乐清市虹桥镇科技园区城东路
生产者名称: 浙江伦特机电有限公司
生产者地址: 浙江省温州市乐清市虹桥镇科技园区城东路
生产企业名称: 浙江伦特机电有限公司
生产企业地址: 浙江省温州市乐清市虹桥镇科技园区城东路

产品名称: 防爆热电阻

系列、规格、型号: WZ 系列

标 准: GB 3836.1-2010、GB 3836.2-2010

上述产品符合强制性产品认证实施规则 CNCA-C23-01:2019 的要求, 特发此证。

发证日期: 2020 年 8 月 6 日

有效期至: 2025 年 8 月 5 日

首次发证日期: 2020 年 8 月 6 日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书的相关信息可通过国家认监委网站 www.cnca.gov.cn 查询



批 准:



上海仪器仪表自控系统检验测试所有限公司



CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



CERTIFICATE NO: 2020322315000498

APPLICANT: Zhejiang Lunte Electromechanical Co., Ltd.
ADDRESS: Chengdong Road, Hongqiao Town, Yueqing City, Zhejiang Province, P.R. China
MANUFACTURER: Zhejiang Lunte Electromechanical Co., Ltd.
ADDRESS: Chengdong Road, Hongqiao Town, Yueqing City, Zhejiang Province, P.R. China
FACTORY: Zhejiang Lunte Electromechanical Co., Ltd.
ADDRESS: Chengdong Road, Hongqiao Town, Yueqing City, Zhejiang Province, P.R. China
PRODUCTNAME: Explosion-proof thermal resistance
SERIES,SPECIFICATION,MODEL: WZ series
STANDARDS: GB 3836.1-2010、GB 3836.2-2010

This is to certify that the above mentioned product(s) complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

Valid from: August 6, 2020

Valid until: August 5, 2025

Date of original certification: August 6, 2020

The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

This certificate is available through CNCA's website: www.cnca.gov.cn



APPROVAL:

Xu JianPing



Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.

<http://www.sitiias.com.cn>

Building 9, 103 Cao Bao Road, Shanghai 200233, China

Tel: +86 21 64510844

S 0000315



中国国家强制性产品认证证书



SITIILAS
Worldwide Access

证书编号: 2020322315000498

附 件

产品名称: 防爆热电阻

型号规格: a-WZPK-b40

a 代表外形结构, 可为 H 或 N

b 代表连接方式, 可为 2, 4, 6, 7 或 9

防爆标志: Ex d IIC T1~T6 Gb

电气参数: U=6.5V

批 准:



上海仪器仪表自控系统检验测试所有限公司