



校准结果

Result of Calibration

证书编号: ZD202112131823

Certificate No.

第 3 页, 共 3 页

Page of

1、外观以及一般性检查:

正常

In view of External and Generality check:

Normal

2、最大称量(Max): 600 g 最小称量(Min): 2 g

Maximum weighting: Minimum weighting:

3、实际分度值(d): 0.01 g 检定分度值(e): 0.1 g

Actual scale: Verification scale:

4、示值误差校准: Calibration of the indication error:

载荷 (g) Weighting	测量误差 (g) Error	允许误差 (g) MPE	结论(Pass/Fail) Conclusion
2	0.00	± 0.05	P
10	0.00	± 0.05	P
50	+0.01	± 0.05	P
100	+0.01	± 0.10	P
200	+0.01	± 0.10	P
500	+0.02	± 0.15	P
600	+0.03	± 0.15	P

5、重复性测试: Repeatability testing

载荷 (g) Weighting	测量误差 (g) Error	允许误差 (g) MPE	结论(Pass/Fail) Conclusion
600	0.01	0.15	P

6、偏载测试: Bias testing

载荷 (g) Weighting	测量误差 (g) Error	允许误差 (g) MPE	结论(Pass/Fail) Conclusion
200	+0.01	± 0.10	P

备注:

Notes:

1 本次测量结果扩展不确定度(Expanded uncertainty of the measurement results):

$U = 0.02 \text{ g} \quad (k=2)$

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)