



校准说明

DIRECTIONS OF CALIBRATION

证书编号: ZD202112131823

第 2 页, 共 3 页

Certificate No.

Page of

1. 本实验室出具的数据均可溯源至国家计量基准。

(All data issued by ZD Test are traced to National Primary Standards.)

2. 本结果仅对当次被测样品有效, 如有疑问请在15个工作日反馈。

(The results is ONLY valid for the tested sample, please feedback to us within 15 working days if you have any question.)

3. 本证书编号具有唯一性, 若替换证书, 自发出后原证书即可作废。

(Each certificate has a unique number. If replaced the certificate, the original certificate will be invalid once the new certificate number is issued.)

4. 证书中如有最大允许误差、判定结果, 仅供参考, 其中 "P" 代表 "Pass", "F" 代表 "Fail".

(In the datasheet, MPE & determination is only for reference, "P" represents "Pass" and "F" represents "Fail".)

5. 本次校准的技术依据及获认可的能力范围, 超出范围的内容未被认可。

(Reference documents and Accredited Scopes for Calibration, Beyond the Scope has not been accredited.):

参照JJG 1036-2008《电子天平检定规程》《Verification Regulation of Electronic Balance》

6. 本次校准使用的主要测量标准 (Main Standards of Measurement Used in the Calibration.):

名称/型号 Description/Model	编号 Serial No.	证书号 Certificate No.	有效期至 Due Date	技术特征 Technique Character
F2等级砝码	0700815	214104495	2022-09-25	F2级合格, 见证书

7. 校准地点、环境条件 (Place and environmental conditions of the calibration)

地点 Place	本公司实验室	温度 Temperature	22 °C	相对湿度 Relative Humidity	55 %
-------------	--------	-------------------	-------	---------------------------	------

8. 建议复校时间间隔:

12 个月, 送检单位可根据实际情况自主决定。

Suggested calibration interval is

12 month or it can be altered depending on the actual usage of the user.