



深圳中电计量测试技术有限公司

Shenzhen Zhongdian Metrology and Testing Technology Co., Ltd

校准证书

CALIBRATION CERTIFICATE



第 1 页, 共 3 页
Page of

证书编号

Certificate No.

ZD202211041521

委托方

Client

合肥市宽虹迎宾楼餐饮管理有限公司

委托方地址

Address

安徽省合肥市蜀山区东流路100号政务中心二区B座

仪器名称

Description

气体流量计

型号规格

Model/Type

JGML-DN

制造厂

Manufacturer

金卡智能集团股份有限公司

出厂编号

Serial No.

031321018481

管理编号

Asset No.

/

样品接收日期

Date of Receipt

2022-11-04

校准日期

Date of Calibration

2022-11-04

批准人 :

Approved Signatory

王小云

审核 :

Inspected by

梁金学

校准 :

Tested by

李



地址 : 广东省深圳市宝安区福永街道白石厦社区东区新开发区4栋301

Add : Room 301, Building 4, New Development Area, East Zone, Baishixia Community, Fuyong Sub-District, Bao'an District, Shenzhen, Guangdong, China

电话 (Tel) : 0755-29888158

邮政编码 (Post Code) : 518103

传真 (Fax) : 0755-29796107



校准说明

DIRECTIONS OF CALIBRATION

证书编号: ZD202211041521

第 2 页, 共 3 页

Certificate No.

Page of

- 1.本实验室出具的数据均可溯源至国家计量基准。
(All data issued by ZD Test are traced to National Primary Standards.)
- 2.本结果仅对当次被测样品有效, 如有疑问请在15个工作日反馈。
(The results is ONLY valid for the tested sample,please feedback to us within 15 working days if you have any question.)
- 3.本证书编号具有唯一性, 若替换证书, 自发出后原证书即可作废。
(Each certificate has a unique number.If replaced the certificate,the original certificate will be invalid once the new certificate number is issued.)
- 4.证书中如有最大允许误差、判定结果, 仅供参考, 其中“P”代表“Pass”, “F”代表“Fail”。
(In the datasheet,MPE & determination is only for reference,"P"represents"Pass"and"F"represents"Fail".)
- 5.本次校准的技术依据及获认可的能力范围, 超出范围的内容未被认可。
(Reference documents and Accredited Scopes for Calibration,Beyond the Scope has not been accredited.):

参照JJG 257-2007《浮子流量计检定规程》《Verification Regulation of Float Meter》

6.本次校准使用的主要测量标准 (Main Standards of Measurement Used in the Calibration.):

名称/型号 Description/Model	编号 Serial No.	证书号 Certificate No.	有效期至 Due Date	技术特征 Technique Character
二等标准水银温度计	2-70	215511200	2023-06-25	$U_{rel}=0.03^{\circ}\text{C}$, $k=2$
流量标准装置	6851452	221284105	2023-03-14	/

7.校准地点、环境条件 (Place and environmental conditions of the calibration)

地点 Place	本公司实验室	温度 Temperature	23.5 °C	相对湿度 Relative Humidity	55 %
-------------	--------	-------------------	---------	---------------------------	------

- 8.建议复校时间间隔 : 12 个月, 送检单位可根据实际情况自主决定。
Suggested calibration interval is 12 month or it can be altered depending on the actual usage of the user.



校准结果

Result of Calibration

证书编号: ZD202211041521

Certificate No.

第 3 页, 共 3 页

Page of

1、外观以及一般性检查:

正常

In view of External and Generality check :

Normal

2、流量校准:

Flow calibration

校准点(m ³ /h)	测量平均值 (m ³ /h)	误差(%)
Calibration point	Average value	Error
0.2	0.2004	+0.20
0.4	0.4007	+0.17
0.8	0.8009	+0.11
1.2	1.2013	+0.11
1.6	1.6018	+0.11

备注:

Notes:

1 本次测量结果扩展不确定度(Expanded uncertainty of the measurement results) :

$$U_{\text{rel}}=0.70\%,k=2$$

2.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)