

计量标准考核证书

Certificate for Examination of Measurement Standard

[2018]鲁东量标考证字第003号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格。特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standard" according to the Law on Metrology of the People's Republic of China.

建标单位名称 山东精益源石油装备检测技术服务有限公司
Possessor of the Measurement Standard

计量标准名称 0.3级测力仪标准装置 代码 12513153
Name of Measurement Standard Code

测量范围 (0.6—300) kN
Measuring range

不确定度或准确度等级
或最大允许误差 0.3级
Uncertainty/Accuracy Class/
Maximum Permissible Error

保存地点 山东精益源石油装备检测技术服务有限公司
Installed in

发证机关 (印章)

The issuing authority

发证日期 2018年8月31日

Date issued

有效期至 2022年8月30日

Date of expiry

0006846

[2018]鲁东量标考证字第008号

计量标准器 Measurement Standard	计量器具名称 Measuring Instrument	型号 Model/Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
	0.3级测力仪标准装置	DD-6-300 kN	6-300 kN	0.3级	西安市帅奇电器有限公司6411
主要配套设备 Main Auxiliary Equipment	温湿度表	WS20280B	(15~45)℃ (15~95)%RH	MPE: ±0.1℃, MPE: 0.2℃ CDP	天津科辉仪表厂 L02
开展的 检定及 校准项目 Verification and Calibration Items	名称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	依据的计量 检定规程或计量技术规范 的编号及名称 Verification Regulation or Technical Specification and Its Code	
	机械式拉力表	(0.6-160) kN	2.0级	JJG455-2000《工作测力仪检定规程》	

考评员：王纯良，