



深圳中电计量测试技术有限公司

Shenzhen Zhongdian Metrology and Testing Technology Co., Ltd

校准证书

CALIBRATION CERTIFICATE



证书编号

Certificate No.

ZD202104250325

第 1 页, 共 3 页

Page of

委托方

Clinet

绵竹市佳伦包装印务有限公司

委托方地址

Address

四川省绵竹市新市场镇

仪器名称

Description

耐破试验机

型号规格

Model/Type

MP-02

制造厂

Manufacturer

济南阿尔法电子有限公司

出厂编号

Serial No.

AF2021022618

管理编号

Asset No.

/

样品接收日期

Date of Receipt

2021-04-25

校准日期

Date of Calibration

2021-04-25

批准人 :

Approved Signatory

王小云

审核 :

Inspected by

巢红

校准 :

Tested by

余小芳



地址 : 广东省深圳市宝安区福永街道白石厦社区东区新开发区4栋301

Add : Room 301, Building 4, New Development Area, East Zone, Baishixia Community, Fuyong Sub-District, Bao'an District, Shenzhen, Guangdong, China

电话 (Tel) : 0755-29888158

邮政编码 (Post Code) : 518103

传真 (Fax) : 0755-29796107



校准说明

DIRECTIONS OF CALIBRATION

证书编号: ZD202104250325

第 2 页, 共 3 页

Certificate No.

Page of

- 1.本实验室出具的数据均可溯源至国家计量基准。
(All data issued by ZD Test are traced to National Primary Standards.)
- 2.本结果仅对当次被测样品有效,如有疑问请在15个工作日反馈。
(The results is ONLY valid for the tested sample, please feedback to us within 15 working days if you have any question.)
- 3.本证书编号具有唯一性,若替换证书,自发出后原证书即可作废。
(Each certificate has a unique number. If replaced the certificate, the original certificate will be invalid once the new certificate number is issued.)
- 4.证书中如有最大允许误差、判定结果,仅供参考,其中“P”代表“Pass”,“F”代表“Fail”。
(In the datasheet, MPE & determination is only for reference, "P" represents "Pass" and "F" represents "Fail".)
- 5.本次校准的技术依据及获认可的能力范围,超出范围的内容未被认可。
(Reference documents and Accredited Scopes for Calibration, Beyond the Scope has not been accredited.):

01《技术说明》

- 6.本次校准使用的主要测量标准 (Main Standards of Measurement Used in the Calibration.):

名称/型号 Description/Model	编号 Serial No.	证书号 Certificate No.	有效期至 Due Date	技术特征 Technique Character
数字压力校验仪	20170306	203207682	2021-10-12	/
标准测力仪	1005	LCL201904048	2021-09-17	0.3级

- 7.校准地点、环境条件 (Place and environmental conditions of the calibration)

地点 Place	委托方现场	温度 Temperature	19 °C	相对湿度 Relative Humidity	58 %
-------------	-------	-------------------	-------	---------------------------	------

- 8.建议复校时间间隔:
Suggested calibration interval is

12 个月, 送检单位可根据实际情况自主决定。
12 month or it can be altered depending on the actual usage of the user.



校准结果

Result of Calibration

证书编号: ZD202104250325

Certificate No.

第 3 页, 共 3 页

Page of

1、外观及各部分相互作用: 正常

In view of externa and the function of each part of interaction: Normal

2、示值校准:

Calibration of the indication

标准值 (kgf/cm ²)	测量平均值 (kgf/cm ²)	不确定度 $U(k=2)$
Reference	Average	Uncertainty
11.61	11.61	0.1%FS

备注:

Notes:

1.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

(以下空白)

(The below is blank)